

	Bezugs-Nr:	HM 8.6
	Stand:	01
Combination Agreement		

COMBINATION AGREEMENT

Alu Rehab AS
Bedriftsveien 23
NO-4353 Klepp Stasjon
Norway

and

AAT Alber Antriebstechnik GmbH
Ehestetter Weg 11
DE-72458 Albstadt
Germany

as manufactures of the

wheelchairs:
Netti III, Netti Svipp, Netti Mini
Netti Svipp Recline, Netti 4U (CE/CED/FA)

product:
"s-max"

hereby declare that the above-mentioned products can be combined under the following conditions;

1. *The necessary fittings must be fitted onto the basic frame of the wheelchair*
Should it prove necessary to place the fittings on other parts of the wheelchair, the specific locations must be determined individually and agreed upon by both parts.
2. *Constructional changes to the wheelchair are allowed.*
Should it prove necessary to add or remove parts, or make any other changes to the wheelchair frame, such modifications must be approved of by the wheelchairs manufacturer.
3. *The assembly instruction in the product manuals must be observed.*
If any special fittings are required for specific wheelchairs, the design and manufacturing thereof must be agreed upon by both parties. The mutually approved fittings must be accompanied by specific assembly instructions.
4. *Technical reservations concerning the combination must be specified.*
Should either of the parties be of the opinion, that a combination requires specific equipment to be fitted onto a wheelchair (e.g. wheels with certain dimensions) this must be agreed upon by both parties and stated in a specific assembly instruction.
5. *The motor unit must not limit any functions or adjustments on the wheelchair.*
Certain changes to the initial adjustment possibilities of the wheelchair are however acceptable, provided they have no influence on safety issues
6. *Any further reservations in relation to the combination must be agreed upon by both parties.*

The motor units mentioned are CE-marked and meet the requirement of Council Directive 93/42/EEC & 2007/47/EEC concerning medical devices of Class 1. Provided that the above conditions are met, the manufacturers hereby declare that the combination of the products also meets these requirements. If second-hand products are to be used for combination, they must first be inspected in accordance with their instructions for recycling. This also applies, if a motor unit is to be fitted onto one of the above-mentioned wheelchairs, which due to its age is not CE-marked.

If regulatory requirements should change or if constructional changes should be made to the products, each of the parties shall evaluate whether or not this affects the validity of the agreement. Each of the parts to the agreement shall notify the other party of any such changes.

Should an incident occur with the combination, which would require notification of national authorities, a copy of the agreement shall be enclosed the notification.

The agreement shall be valid for an indefinite period of time as from its execution. However, it can be terminated by either party in writing by giving 30 days notice.

AAT Alber Antriebstechnik GmbH



Ebingen · Ehstetter Weg 11
D-72458 Albstadt
Tel +49.74 31.12 95 - 0
Fax +49.74 31.12 95 - 35

20.01.12 Leiv Egil Lonnu

Date

Alu Rehab AS

20.01.12 L. Bruno

Date

Luigi Bruno
AAT Alber Antriebstechnik GmbH

Formular erstellt am: 15.02.10	erstellt von: L. Bruno	geprüft am: 15.02.10	geprüft von: L. Bruno	Seite 1 von 1
HM 8.6 Combination Agreement				